

REF	611-2	6×1mL	EXP	2021-10-26	LOT	3A4A571	更新日期：	2019-08
								1 / 1

		批号：3A4A571	
项目\仪器\试剂\方法	单位	均值	+ / - 2 SD
(HbA1c)糖化血红蛋白A1c			
Bio-Rad D-10/100 Dedicated \ High performance liquid chromatography (HPLC)	%	11.4	9.2-13.6
Medical MQ-2000PT Dedicated \ High performance liquid chromatography (HPLC)	%	11.3	9.1-13.5
Premier Ultra2 Dedicated \ High performance liquid chromatography (HPLC)	%	12.1	10.3-13.9
Roche Hitachi/Hitachi Systems Dedicated \ immunoturbidimetric	%	10.8	9.0-12.6
Tosoh G7/G8 series Dedicated \ High performance liquid chromatography (HPLC)	%	10.9	8.7-13.1

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。