



REF	671-3	6×1mL	EXP	2026-06-03	LOT	4005B98	更新日期：	2023-12
								1 / 2

本报告含以下项目的示范实验室数据

参数 (Parameters)	参数 (Parameters)	参数 (Parameters)
(APTT)活化部分凝血活酶时间	(AT-III)抗凝血酶III	(D-D(DDU)) D-二聚体(DDU)
(D-D(FEU)) D-二聚体(FEU)	(FIB)纤维蛋白原	(PT)血浆凝血酶原时间

注：此报告所提供的项目与数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来。此报告不可代替产品说明书。此报告仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。

REF	671-3	6×1mL	EXP	2026-06-03	LOT	4005B98	更新日期：	2023-12
								2 / 2

批号：4005B98			
项目\仪器\试剂方法	单位	均值	+ / - 2 SD
(APTT)活化部分凝血活酶时间			
Werfen ACL series			
Werfen Hemosil SynthAsil APTT \ Light scattering	Seconds	103	82.4-124
Werfen ACL series			
Werfen Hemosil APTT-SP \ Light scattering	Seconds	95.8	76.6-115
Stago STA series			
Stago \ Magnetic Beads	Seconds	86.6	69-104
Sysmex CA 1500/6000/7000			
Siemens Dade Actin(APTT) \ Light scattering	Seconds	146	117-175
Sysmex CA 1500/6000/7000			
Siemens Pathrombin SL(APTT) \ Light scattering	Seconds	151	121-181
(AT-III)抗凝血酶III			
Werfen ACL series			
Werfen Hemosil Liquid Antithrombin (AT□) \ Nephelometry	%	25.5	20.3-30.7
Sysmex CA 1500/6000/7000			
Siemens Berichrom Antithrombin III (AT□) \ Nephelometry	%	27.4	21.9-32.9
(D-D(DDU)) D-二聚体(DDU)			
Werfen ACL series			
Werfen Hemosil D-Dimer HS \ Nephelometry	mg/L	1.14	0.912-1.37
Werfen ACL series			
Werfen Hemosil D-Dimer \ Nephelometry	mg/L	1.09	0.87-1.31
(D-D(FEU)) D-二聚体(FEU)			
Sysmex CA 1500/6000/7000			
Siemens Innovance Antithrombin (AT□) \ Nephelometry	mg/L	3.74	2.99-4.49
(FIB)纤维蛋白原			
Werfen ACL series			
Werfen Hemosil QFA Thrombin (FIB) \ Light scattering	g/L	1.24	0.99-1.49
Werfen ACL series			
Werfen Hemosil Fib-C \ Light scattering	g/L	1.13	0.9-1.36
Stago STA series			
Stago \ Magnetic Beads	g/L	1.38	1.1-1.66
Sysmex CA 1500/6000/7000			
Siemens Multifibren U (FIB) \ Light scattering	g/L	1.78	1.42-2.14
(PT)血浆凝血酶原时间			
Werfen ACL series			
Werfen Hemosil PT Recombiplastin \ Light scattering	Seconds	48.8	38.8-58.8
Werfen ACL series			
Werfen Hemosil PT Recombiplastin \ Light scattering	INR	4.47	3.53-5.41
Werfen ACL series			
Werfen Hemosil PT Recombiplastin 2G \ Light scattering	Seconds	53.1	42.3-63.9
Werfen ACL series			
Werfen Hemosil PT Recombiplastin 2G \ Light scattering	INR	4.52	3.6-5.44
Werfen ACL series			
Werfen Hemosil PT-Fib HS Plus \ Light scattering	Seconds	63.9	50.9-76.9
Werfen ACL series			
Werfen Hemosil PT-Fib HS Plus \ Light scattering	INR	5.19	4.15-6.23
Stago STA series			
Stago \ Magnetic Beads	Seconds	52.9	42.3-63.5
Stago STA series			
Stago \ Magnetic Beads	INR	5.91	4.71-7.11
Sysmex CA 1500/6000/7000			
Siemens PT Innovin \ Light scattering	Seconds	43.6	34.6-52.6
Sysmex CA 1500/6000/7000			
Siemens PT Innovin \ Light scattering	INR	4.12	3.3-4.94
Sysmex CA 1500/6000/7000			
Siemens Thromborel S (PT) \ Light scattering	Seconds	50.5	40.3-60.7
Sysmex CA 1500/6000/7000			
Siemens Thromborel S (PT) \ Light scattering	INR	4.64	3.64-5.64

注：此报告所提供的项目与数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来。此报告不可代替产品说明书。此报告仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。