

REF 671-1	6×1mL	EXP 2021-12-11	LOT 4000BE2	更新日期： 2019-09
				1 / 1

批号：4000BE2			
项目\仪器\试剂\方法	单位	均值	+ / - 2 SD
(APTT)活化部分凝血活酶时间			
IL ACL series			
IL \ Light scattering	Seconds	32.6	26.2-39.0
Stago STA series			
Dedicated \ Magnetic Beads	Seconds	35.2	28.8-41.6
Sysmex CA 1500/6000/7000			
Siemens \ Light scattering	Seconds	33.8	27.4-40.2
Sysmex CS 5100			
Siemens \ Light scattering	Seconds	34.2	27.8-40.6
(ATIII)抗凝血酶III			
IL ACL series			
Dedicated \ Nephelometry	%	95.9	76.7-115.1
(D-D(FEU)) D-二聚体(FEU)			
Stago STA series			
Dedicated \ Latex agglutination (LA)	mg/L	1.48	1.18-1.78
Sysmex CA 1500/6000/7000			
SekiSui \ Nephelometry	mg/L	1.55	1.27-1.83
Sysmex CS 5100			
Dedicated \ Nephelometry	mg/L	1.58	1.30-1.86
(FDP)纤维蛋白(原)降解产物			
IL ACL series			
SekiSui \ Nephelometry	mg/L	4.32	3.46-5.18
Stago STA series			
Dedicated \ Nephelometry	mg/L	4.82	3.94-5.70
Sysmex CA 1500/6000/7000			
SekiSui \ Nephelometry	mg/L	5.06	4.08-6.04
(FIB)纤维蛋白原			
IL ACL series			
IL \ Light scattering	g/L	3.54	2.82-4.26
Stago STA series			
Dedicated \ Magnetic Beads	g/L	3.62	2.90-4.34
Sysmex CA 1500/6000/7000			
Siemens \ Light scattering	g/L	3.22	2.58-3.86
Sysmex CS 5100			
Siemens \ Light scattering	mg/dL	306	244-368
(PT)血浆凝血酶原时间			
IL ACL series			
IL \ Light scattering	Seconds	11.6	9.2-14.0
Stago STA series			
Dedicated \ Magnetic Beads	Seconds	13.1	10.7-15.5
Sysmex CA 1500/6000/7000			
Siemens \ Light scattering	Seconds	12.5	10.1-14.9
Sysmex CS 5100			
Siemens \ Light scattering	Seconds	12.7	10.3-15.1
(TT)凝血酶时间			
IL ACL series			
SUNBIO \ Light scattering	Seconds	16.8	13.2-20.4
Stago STA series			
Dedicated \ Magnetic Beads	Seconds	17.8	14.2-21.4
Sysmex CA 1500/6000/7000			
Dedicated \ Light scattering	Seconds	18.6	15.0-22.2

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。